

T:SLIM X2 INSULIN PUMP TRAINING

Supply Checklist







t:slim X2 Insulin Pump Supplies



Additional Supplies



with test strips



*Glucose monitoring supplies sold separately.

Important Safety Information: RX ONLY. The t:slim X2 pump with interoperable technology (the pump) and Control-IQ+ technology (Control-IQ+) are intended for single patient use. The pump and Control-IQ+ are indicated for use with NovoLog or Humalog U-100 insulin. t:slim X2 insulin pump: The pump is intended for the subcutaneous delivery of insulin, at set and variable rates, for the management of diabetes mellitus in persons requiring insulin. The pump is able to reliably and securely communicate with compatible, digitally connected devices, including automated insulin dosing software, to receive, execute, and confirm commands from these devices. The pump is indicated for use in persons 2 years of age and greater. Control-IQ+ technology: Control-IQ+ technology is intended for use with compatible integrated continuous glucose monitors (iCGM, sold separately) and alternate controller enabled (ACE) pumps to automatically increase, decrease, and suspend delivery of basal insulin based on iCGM readings and predicted glucose values. It can also deliver correction boluses when the glucose value is predicted to exceed a predefined threshold. Control-IQ+ technology is intended for the management of type 1 diabetes mellitus in persons 2 years of age and greater and of type 2 diabetes mellitus in persons 18 years of age and greater.

WARNING: Control-IQ+ technology should not be used in anyone under the age of 2 years old with type 1 diabetes or under the age of 18 years old with type 2 diabetes. It should also not be used in patients who require less than a total daily insulin dose of 5 units of insulin per day or who weigh less than 20 pounds (9 kilograms), as those are the required minimum values needed for Control-IQ+ to operate safely.

Users of the pump and Control-IQ+ must: use the insulin pump, iCGM, and all other system components in accordance with their respective instructions for use. Failure to follow these instructions for use could result in an over delivery or under delivery of insulin. This can cause hypoglycemia (low BG) or hyperglycemia (high BG) events. Visit tandemdiabetes.com/safetyinfo for additional important safety information.

© 2025 Tandem Diabetes Care, Inc. All rights reserved. Tandem Diabetes Care, Tandem logos, Control-IQ+, and t:slim X2 are either registered trademarks or trademarks of Tandem Diabetes Care, Inc. in the United States and/or other countries. Dexcom, Dexcom G6, Dexcom G7, and any related logos and design marks are either registered trademarks or trademarks of Dexcom, Inc. in the United States and/or other countries. The sensor housing, FreeStyle, Libre, and related brand marks are marks of Abbott and used with permission. All third-party marks are the property of their respective owners. MAT-0315_1



